



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong
Tel: (852) 2333 6823 Fax: (852) 2333 1316

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

Sampling Date	Job No.	Unit	Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
			MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
			mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
06/12/2022	R221930		0.22	0.15	4.64	4.83	-4.0	10	94.2	Pass
08/12/2022	R221940		0.22	0.07	5.13	5.41	-5.3	10	94.9	Pass
10/12/2022	R221946		0.22	0.08	4.94	5.20	-5.1	10	94.9	Pass
13/12/2022	R221970		0.22	0.11	5.38	5.58	-3.6	10	96.6	Pass
15/12/2022	R221980		0.22	0.05	4.47	4.32	3.4	10	97.6	Pass
17/12/2022	R221981		0.22	0.09	5.34	5.69	-6.4	10	94.7	Pass
20/12/2022	R221999		0.22	0.10	4.98	4.72	5.4	10	94.6	Pass
22/12/2022	R222000		0.22	0.08	4.38	4.55	-3.8	10	95.8	Pass
24/12/2022	R222025		0.22	0.07	4.90	4.81	1.8	10	92.1	Pass
26/12/2022	R222076		0.22	0.05	4.67	4.53	3.0	10	97.0	Pass
28/12/2022	R222077		0.22	0.09	4.76	4.94	-3.7	10	96.0	Pass
30/12/2022	R222089		0.22	0.08	4.90	4.83	1.4	10	98.9	Pass