



# Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong  
 Tel: (852) 2333 6823 Fax: (852) 2333 1316

## Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
02/01/2023	R230006	0.22	0.08	4.53	4.40	2.9	10	94.5	Pass
04/01/2023	R230013	0.22	0.11	4.53	4.33	4.5	10	95.6	Pass
06/01/2023	R230014	0.22	0.09	6.87	6.40	7.1	10	95.1	Pass
10/01/2023	R230042	0.22	0.05	5.89	6.02	-2.2	10	93.1	Pass
12/01/2023	R230052	0.22	0.06	5.03	4.88	3.0	10	95.8	Pass
14/01/2023	R230055	0.22	0.08	4.46	4.73	-5.9	10	92.1	Pass
17/01/2023	R230060	0.22	0.07	5.19	5.44	-4.7	10	94.8	Pass
19/01/2023	R230064	0.22	0.07	5.35	5.22	2.5	10	95.8	Pass
21/01/2023	R230118	0.22	0.08	4.84	4.53	6.6	10	92.2	Pass
25/01/2023	R230127	0.22	0.07	5.29	5.15	2.7	10	95.0	Pass
27/01/2023	R230130	0.22	0.06	5.87	5.66	3.6	10	96.8	Pass
30/01/2023	R230142	0.22	0.09	5.65	6.02	-6.3	10	94.8	Pass