



# Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong  
 Tel: (852) 2333 6823 Fax: (852) 2333 1316

## Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
01/03/2023	R230328	0.22	0.08	4.60	4.85	-5.3	10	94.6	Pass
03/03/2023	R230329	0.22	0.09	4.88	5.04	-3.2	10	96.3	Pass
07/03/2023	R230335	0.22	0.09	5.74	5.56	3.2	10	95.7	Pass
09/03/2023	R230336	0.22	0.08	4.91	5.07	-3.2	10	95.0	Pass
11/03/2023	R230359	0.22	0.10	5.34	5.63	-5.3	10	97.1	Pass
13/03/2023	R230360	0.22	0.08	4.77	5.07	-6.1	10	96.4	Pass
15/03/2023	R230384	0.22	0.11	4.80	4.91	-2.3	10	92.9	Pass
17/03/2023	R230385	0.22	0.06	6.34	6.51	-2.7	10	96.8	Pass
21/03/2023	R230428	0.22	0.06	6.02	6.30	-4.6	10	94.2	Pass
23/03/2023	R230429	0.22	0.07	5.01	4.82	3.9	10	95.0	Pass
25/03/2023	R230433	0.22	0.07	6.29	6.42	-2.1	10	93.5	Pass
28/03/2023	R230434	0.22	0.08	4.77	4.59	3.9	10	97.6	Pass
30/03/2023	R230472	0.22	0.08	5.66	5.45	3.8	10	92.7	Pass