



# Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong  
 Tel: (852) 2333 6823 Fax: (852) 2333 1316

## Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
01/04/2023	R230473	0.22	0.11	4.45	4.59	-3.1	10	96.4	Pass
04/04/2023	R230498	0.22	0.09	4.12	4.42	-7.0	10	92.7	Pass
06/04/2023	R230499	0.22	0.09	4.61	4.88	-5.7	10	97.8	Pass
08/04/2023	R230500	0.22	0.05	3.90	4.12	-5.5	10	92.6	Pass
11/04/2023	R230535	0.22	0.06	5.65	5.34	5.6	10	92.7	Pass
13/04/2023	R230536	0.22	0.07	5.05	4.86	3.8	10	93.0	Pass
15/04/2023	R230537	0.22	0.10	5.29	5.08	4.0	10	97.0	Pass
17/04/2023	R230571	0.22	0.08	5.44	5.67	-4.1	10	96.2	Pass
19/04/2023	R230601	0.22	0.07	5.03	4.85	3.6	10	93.1	Pass
21/04/2023	R230602	0.22	0.06	6.19	6.39	-3.2	10	93.6	Pass
25/04/2023	R230603	0.22	0.08	4.97	5.28	-6.0	10	91.6	Pass
27/04/2023	R230604	0.22	0.07	4.93	4.79	2.9	10	95.3	Pass
29/04/2023	R230621	0.22	0.10	5.72	5.47	4.5	10	93.5	Pass