



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong
Tel: (852) 2333 6823 Fax: (852) 2333 1316

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
01/07/2023	R230933	0.22	0.07	4.34	4.15	4.5	10	94.5	Pass
04/07/2023	R230934	0.22	0.06	5.02	4.79	4.7	10	91.3	Pass
06/07/2023	R230954	0.22	0.07	4.91	5.27	-7.1	10	94.6	Pass
08/07/2023	R230955	0.22	0.09	5.38	5.05	6.3	10	92.8	Pass
11/07/2023	R230974	0.22	0.06	4.09	3.89	5.0	10	95.3	Pass
13/07/2023	R230975	0.22	0.05	5.12	5.40	-5.3	10	95.0	Pass
15/07/2023	R231006	0.22	0.07	4.00	3.88	3.0	10	93.3	Pass
18/07/2023	R231007	0.22	0.06	4.53	4.39	3.1	10	96.5	Pass
19/07/2023	R231037	0.22	0.09	4.12	4.36	-5.7	10	91.3	Pass
20/07/2023	R231038	0.22	0.08	5.97	5.77	3.4	10	93.8	Pass
22/07/2023	R231078	0.22	0.09	4.89	5.13	-4.8	10	92.8	Pass
24/07/2023	R231079	0.22	0.06	3.42	3.20	6.6	10	91.3	Pass
26/07/2023	R231080	0.22	0.07	5.72	5.50	3.9	10	95.5	Pass
28/07/2023	R231104	0.22	0.08	5.16	5.37	-4.0	10	95.6	Pass
31/07/2023	R231105	0.22	0.07	4.28	4.45	-3.9	10	92.9	Pass