



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong
 Tel: (852) 2333 6823 Fax: (852) 2333 1316

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. / Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
R231381	04/09/2023	0.22	0.07	5.55	5.18	6.9	10	93.2	Pass
R231382	06/09/2023	0.22	0.06	6.17	6.40	-3.7	10	92.6	Pass
R231383	09/09/2023	0.22	0.09	4.50	4.37	2.9	10	90.9	Pass
R231384	11/09/2023	0.22	0.06	4.25	4.51	-5.9	10	93.6	Pass
R231460	13/09/2023	0.22	0.08	4.31	4.16	3.5	10	91.4	Pass
R231461	15/09/2023	0.22	0.06	4.79	4.91	-2.5	10	95.7	Pass
R231488	19/09/2023	0.22	0.07	4.89	5.11	-4.4	10	93.7	Pass
R231489	21/09/2023	0.22	0.07	5.46	5.26	3.7	10	93.2	Pass
R231490	23/09/2023	0.22	0.05	5.43	5.24	3.6	10	91.7	Pass
R231491	25/09/2023	0.22	0.06	5.12	5.45	-6.2	10	94.4	Pass
R231535	27/09/2023	0.22	0.09	4.86	5.03	-3.4	10	93.8	Pass
R231563	29/09/2023	0.22	0.06	5.83	5.66	3.0	10	92.7	Pass