



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong
 Tel: (852) 2333 6823 Fax: (852) 2333 1316

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

| | | Method Blank Report | | Duplicate Report | | | Sample Spike Report | | Pass / Fail |
|---------------|--------------|---------------------|--------|------------------|------------------|------|---------------------|----------------|-------------|
| | | MDL | Result | Original Result | Duplicate Result | RPD | Spike concentration | Spike Recovery | |
| Sampling Date | Job No. Unit | mg/L | mg/L | mg/L | mg/L | % | mg/L | % | / |
| 02/11/2023 | R231770 | 0.22 | 0.07 | 4.41 | 4.67 | -5.7 | 10 | 91.1 | Pass |
| 04/11/2023 | R231806 | 0.22 | 0.08 | 4.22 | 4.05 | 4.1 | 10 | 95.3 | Pass |
| 07/11/2023 | R231807 | 0.22 | 0.11 | 5.05 | 5.19 | -2.7 | 10 | 94.2 | Pass |
| 09/11/2023 | R231808 | 0.22 | 0.10 | 4.03 | 4.25 | -5.3 | 10 | 92.0 | Pass |
| 11/11/2023 | R231809 | 0.22 | 0.07 | 4.54 | 4.31 | -4.2 | 10 | 93.7 | Pass |
| 14/11/2023 | R231838 | 0.22 | 0.08 | 5.48 | 5.36 | 2.2 | 10 | 91.8 | Pass |
| 16/11/2023 | R231839 | 0.22 | 0.09 | 5.67 | 5.51 | 2.9 | 10 | 93.2 | Pass |
| 18/11/2023 | R231840 | 0.22 | 0.07 | 5.50 | 5.71 | -3.8 | 10 | 92.8 | Pass |
| 21/11/2023 | R231841 | 0.22 | 0.09 | 5.14 | 5.34 | -3.8 | 10 | 93.0 | Pass |
| 23/11/2023 | R231842 | 0.22 | 0.07 | 4.82 | 5.10 | -5.7 | 10 | 94.0 | Pass |
| 25/11/2023 | R231931 | 0.22 | 0.08 | 4.70 | 4.47 | 5.0 | 10 | 91.8 | Pass |
| 28/11/2023 | R231932 | 0.22 | 0.09 | 4.27 | 4.08 | 4.6 | 10 | 93.0 | Pass |
| 30/11/2023 | R231944 | 0.22 | 0.10 | 4.18 | 4.33 | -3.5 | 10 | 91.3 | Pass |