



# Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong  
Tel: (852) 2333 6823 Fax: (852) 2333 1316

## Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

Sampling Date	Job No.	Unit	Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
			MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
			mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
02/01/2024	R240133		0.22	0.06	3.79	3.90	-2.9	10	95.9	Pass
04/01/2024	R240134		0.22	0.10	4.48	4.55	-1.6	10	92.9	Pass
06/01/2024	R240185		0.22	0.07	4.30	4.38	-1.8	10	91.3	Pass
09/01/2024	R240186		0.22	0.10	3.77	3.85	-2.1	10	90.2	Pass
11/01/2024	R240187		0.22	0.07	3.91	3.74	4.4	10	91.9	Pass
13/01/2024	R240211		0.22	0.09	4.09	4.24	-3.6	10	95.7	Pass
16/01/2024	R240212		0.22	0.06	4.03	4.16	-3.2	10	93.6	Pass
18/01/2024	R240247		0.22	0.06	3.23	3.16	2.2	10	94.7	Pass
20/01/2024	R240282		0.22	0.10	4.20	4.32	-2.8	10	93.0	Pass
22/01/2024	R240283		0.22	0.07	3.83	3.99	-4.1	10	95.2	Pass
23/01/2024	R240302		0.22	0.08	3.10	3.20	-3.2	10	94.3	Pass
25/01/2024	R240303		0.22	0.09	4.30	4.19	2.6	10	92.4	Pass
27/01/2024	R240344		0.22	0.08	5.71	5.58	2.3	10	93.0	Pass
30/01/2024	R240345		0.22	0.08	4.14	3.97	4.2	10	93.1	Pass