



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong
Tel: (852) 2333 6823 Fax: (852) 2333 1316

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
01/02/2024	R240384	0.22	0.09	3.86	3.67	5.0	10	90.8	Pass
03/02/2024	R240385	0.22	0.08	3.52	3.38	4.1	10	95.6	Pass
05/02/2024	R240399	0.22	0.12	3.89	3.99	-2.5	10	92.2	Pass
07/02/2024	R240400	0.22	0.10	3.97	3.89	2.0	10	92.7	Pass
15/02/2024	R240457	0.22	0.09	4.39	4.20	4.4	10	94.5	Pass
17/02/2024	R240459	0.22	0.11	3.81	3.92	-2.8	10	94.9	Pass
19/02/2024	R240460	0.22	0.09	4.07	4.18	-2.7	10	92.8	Pass
21/02/2024	R240461	0.22	0.07	4.45	4.28	3.9	10	93.0	Pass
23/02/2024	R240531	0.22	0.11	3.93	3.75	4.7	10	91.4	Pass
27/02/2024	R240551	0.22	0.09	4.73	4.91	-3.7	10	93.5	Pass
29/02/2024	R240552	0.22	0.08	3.28	3.41	-3.9	10	90.8	Pass