



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong
Tel: (852) 2333 6823 Fax: (852) 2333 1316

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

Sampling Date	Job No.	Unit	Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
			MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
			mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
02/03/2024	R240576		0.22	0.09	3.56	3.66	-2.8	10	96.2	Pass
05/03/2024	R240577		0.22	0.07	4.52	4.30	5.0	10	92.7	Pass
07/03/2024	R240608		0.22	0.07	3.84	3.77	1.8	10	93.5	Pass
09/03/2024	R240609		0.22	0.08	3.43	3.52	-2.6	10	96.8	Pass
12/03/2024	R240643		0.22	0.10	3.95	4.09	-3.5	10	93.6	Pass
14/03/2024	R240644		0.22	0.09	4.03	4.14	-2.7	10	95.3	Pass
16/03/2024	R240647		0.22	0.08	4.95	4.81	2.9	10	92.9	Pass
18/03/2024	R240648		0.22	0.11	3.53	3.76	-6.3	10	95.8	Pass
20/03/2024	R240696		0.22	0.09	4.08	3.86	5.5	10	94.8	Pass
22/03/2024	R240730		0.22	0.09	4.52	4.41	2.5	10	91.6	Pass
25/03/2024	R240774		0.22	0.08	4.91	5.09	-3.6	10	95.3	Pass
27/03/2024	R240775		0.22	0.08	3.70	3.89	-5.0	10	94.9	Pass
29/03/2024	R240776		0.22	0.07	3.54	3.40	4.6	10	93.0	Pass