



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong
 Tel: (852) 2333 6823 Fax: (852) 2333 1316

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
02/04/2024	R240837	0.22	0.07	3.79	3.65	3.8	10	94.3	Pass
04/04/2024	R240838	0.22	0.08	4.05	4.25	-4.8	10	93.1	Pass
06/04/2024	R240839	0.22	0.07	4.31	4.11	4.8	10	93.0	Pass
09/04/2024	R240840	0.22	0.10	4.17	4.32	-3.5	10	93.6	Pass
11/04/2024	R240841	0.22	0.08	3.71	3.87	-4.2	10	92.6	Pass
13/04/2024	R240866	0.22	0.10	3.93	4.18	-6.2	10	91.5	Pass
15/04/2024	R240867	0.22	0.09	4.92	4.74	3.7	10	93.9	Pass
17/04/2024	R240868	0.22	0.08	4.64	4.40	5.3	10	91.8	Pass
19/04/2024	R240889	0.22	0.10	4.54	4.34	2.4	10	94.9	Pass
23/04/2024	R240916	0.22	0.11	3.69	3.88	-5.0	10	92.7	Pass
25/04/2024	R240954	0.22	0.11	4.05	3.83	5.6	10	92.0	Pass
27/04/2024	R240955	0.22	0.07	4.33	4.53	-4.5	10	94.5	Pass
30/04/2024	R240967	0.22	0.08	5.10	5.34	-4.6	10	96.1	Pass