



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong
Tel: (852) 2333 6823 Fax: (852) 2333 1316

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
02/05/2024	R241010	0.22	0.07	5.11	4.97	2.8	10	94.3	Pass
06/05/2024	R241053	0.22	0.08	4.30	4.44	-3.2	10	90.7	Pass
08/05/2024	R241054	0.22	0.07	5.36	5.06	5.8	10	91.3	Pass
10/05/2024	R241055	0.22	0.11	4.89	4.68	4.4	10	91.8	Pass
14/05/2024	R241056	0.22	0.09	5.72	5.86	-2.4	10	94.6	Pass
16/05/2024	R241057	0.22	0.10	4.29	4.54	-5.7	10	92.6	Pass
18/05/2024	R241104	0.22	0.10	4.70	4.48	4.8	10	93.4	Pass
20/05/2024	R241105	0.22	0.08	4.02	4.21	-4.6	10	92.6	Pass
22/05/2024	R241106	0.22	0.07	4.21	4.26	-1.2	10	92.3	Pass
25/05/2024	R241107	0.22	0.11	4.67	4.86	-4.0	10	91.8	Pass
27/05/2024	R241158	0.22	0.09	5.11	5.28	-3.3	10	91.6	Pass
29/05/2024	R241159	0.22	0.10	4.13	3.90	5.7	10	92.1	Pass
31/05/2024	R241225	0.22	0.08	4.05	4.21	-3.9	10	94.2	Pass