



Acumen Laboratory and Testing Limited

Workshop 04, 7/F, The Whitney, No. 183 Wai Yip Street, Kwun Tong, Kowloon

Tel: (852) 2333 6823 Fax: (852) 2333 1316

Page 1 of 1

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
02/07/2025	R251547	0.22	0.10	4.77	4.92	-3.10	10	92.6	Pass
04/07/2025	R251568	0.22	0.09	5.05	4.89	3.22	10	92.7	Pass
07/07/2025	R251578	0.22	0.09	3.32	3.40	-2.38	10	94.3	Pass
09/07/2025	R251611	0.22	0.07	3.62	3.49	3.66	10	93.7	Pass
12/07/2025	R251630	0.22	0.08	5.04	5.20	-3.13	10	93.5	Pass
14/07/2025	R251656	0.22	0.11	3.20	3.11	2.85	10	94.9	Pass
16/07/2025	R251679	0.22	0.07	5.13	5.32	-3.64	10	93.3	Pass
18/07/2025	R251688	0.22	0.10	3.65	3.50	4.20	10	93.7	Pass
21/07/2025	R251690	0.22	0.11	4.65	4.86	-4.42	10	94.5	Pass
23/07/2025	R251728	0.22	0.10	4.75	4.61	2.99	10	92.9	Pass
25/07/2025	R251746	0.22	0.09	4.54	4.69	-3.25	10	93.3	Pass
28/07/2025	R251749	0.22	0.10	4.23	4.07	3.86	10	94.8	Pass
30/07/2025	R251781	0.22	0.09	3.50	3.59	-2.54	10	93.6	Pass