

專業化驗有限公司

QUALITY PRO TEST-CONSULT LIMITED

Unit 10, 5/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong

Email: info@qualityprotest.com; Website: www.qualityprotest.com

Tel: (852) 3956 8717; Fax: (852) 3956 3928

## REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Test Report No. : R-BE090186  
Date of Issue : 26 September 2025  
Page No. : 1 of 2

### PART A - CUSTOMER INFORMATION

Acuity Sustainability Consulting Limited

Unit 1608, 16/F, Tower B, Manulife Fin. Centre 223 - 231 Wai Yip Street, Kwun Tong,  
Kowloon (HK) Hong Kong

### PART B - SAMPLE INFORMATION

Name of Equipment : YSI ProDSS (Multi Parameters)  
Manufacturer : YSI  
Serial Number : 22D100436  
Date of Received : 19 September 2025  
Date of Calibration : 25 September 2025  
Date of Next Calibration : 24 December 2025  
Request No. : D-BE090186

### PART C - REFERENCE METHODS/ DOCUMENTS FOR THE CALIBRATION

Test Parameter	Reference Method
pH value	APHA 21e 4500-H <sup>+</sup> B
Temperature	Section 6 of international Accreditation New Zealand Technical Guide no. 3 Second edition March 2008: Working Thermometer Calibration Procedure
Dissolved oxygen	APHA 23e 4500-O G (Membrane Electrode Method)
Conductivity	APHA 21e 2510 B
Salinity	APHA 21e 2520 B
Turbidity	APHA 21e 2130 B (Nephelometric Method)

### PART D - CALIBRATION RESULT

#### (1) pH value

Target (pH unit)	Display Reading (pH unit)	Tolerance (pH unit)	Result
4.00	4.00	0	Satisfactory
7.42	7.42	0	Satisfactory
10.01	10.00	-0.01	Satisfactory

Tolerance of pH value should be less than  $\pm 0.2$  (pH unit)

#### (2) Temperature

Reading of Ref. thermometer (°C)	Display Reading (°C)	Tolerance (°C)	Result
35.6	35.0	-0.6	Satisfactory
24.9	24.6	-0.3	Satisfactory
11.8	11.7	-0.1	Satisfactory

Tolerance of Temperature should be less than  $\pm 2.0$  (°C)

#### (3) Dissolved oxygen


Expected Reading (mg/L)	Display Reading (mg/L)	Tolerance (mg/L)	Result
7.71	7.89	0.18	Satisfactory
6.52	6.50	-0.02	Satisfactory
3.14	3.32	0.18	Satisfactory
0.02	0.08	0.06	Satisfactory

Tolerance of Dissolved oxygen should be less than  $\pm 0.5$  (mg/L)

--- CONTINUED ON NEXT PAGE ---

AUTHORIZED  
SIGNATORY:

FUNG Yuen-ching  
Laboratory Manager



專業化驗有限公司

QUALITY PRO TEST-CONSULT LIMITED

Unit 10, 5/F, Wah Wai Centre, 38-40 Au Pui Wan St., Fotan, Hong Kong

Email: info@qualityprotest.com; Website: www.qualityprotest.com

Tel: (852) 3956 8717; Fax: (852) 3956 3928

## REPORT OF EQUIPMENT PERFORMANCE CHECK/ CALIBRATION

Test Report No. : R-BE090186

Date of Issue : 26 September 2025

Page No. : 2 of 2

### (4) Conductivity

Expected Reading ( $\mu\text{S/cm}$ at 25°C )	Display Reading ( $\mu\text{S/cm}$ at 25°C )	Tolerance ( % )	Result
146.9	160.1	9.0	Satisfactory
1412	1426	1.0	Satisfactory
12890	11741	-8.9	Satisfactory
58670	53841	-8.2	Satisfactory
111900	104478	-6.6	Satisfactory

Tolerance of Conductivity should be less than  $\pm 10.0$  ( % )

### (5) Salinity

Expected Reading ( g/L )	Display Reading ( g/L )	Tolerance ( % )	Result
10	9.69	-3.10	Satisfactory
20	18.93	-5.35	Satisfactory
30	30.79	2.63	Satisfactory

Tolerance of Salinity should be less than  $\pm 10.0$  ( % )

### (6) Turbidity

Expected Reading ( NTU )	Display Reading ( NTU )	Tolerance <sup>(a)</sup> ( % )	Result
0	0.36	-	Satisfactory
10	9.46	-5.4	Satisfactory
20	18.47	-7.7	Satisfactory
100	94.00	-6.0	Satisfactory
800	734.12	-8.2	Satisfactory

Tolerance of Turbidity should be less than  $\pm 10.0$  ( % )

<sup>(a)</sup> For 0 NTU, Display Reading should be less than 1 NTU

#### Remark(s):

1. The "Date of Next Calibration" is set in accordance with the best practices of QPT or relevant international standards.
2. The calibration results apply exclusively to the equipment as received.
3. This report verifies the equipment's performance by testing it with independent reference material and comparing the results against a calibrated secondary standard.
4. "Displayed Reading" refers to the value shown on the unit under calibration, irrespective of its internal precision or significant figures.
5. The "Tolerance Limit" represents the acceptance criteria used by Quality Pro Test-Consult Ltd. for similar equipment, as derived from relevant international standards.

--- END OF REPORT ---



ALS Technichem (HK) Pty Ltd

11/F., Chung Shun Knitting Centre,  
1 - 3 Wing Yip Street,  
Kwai Chung, N.T., Hong Kong

T: +852 2610 1044

F: +852 2610 2021

www.alsglobal.com

## **REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION**

**CONTACT:** JOE HO  
**CLIENT:** AURECON HONG KONG LIMITED  
**ADDRESS:** UNIT 1608, 16/F, TOWER B,  
MANULIFE FINANCIAL CENTRE,  
223-231 WAI YIP STREET,  
KWUN TONG, HONG KONG

**WORK ORDER:** HK2543901  
**SUB-BATCH:** 0  
**LABORATORY:** HONG KONG  
**DATE RECEIVED:** 11-Oct-2025  
**DATE OF ISSUE:** 21-Oct-2025

### GENERAL COMMENTS

The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

The "Tolerance Limit" quoted is the acceptance criteria applicable for similar equipment used by the laboratory or quoted from relevant international standards.

The "Next Calibration Date" is recommended according to best practice principle as practised by the laboratory or quoted from relevant international standards.

The validity of equipment/ meter performance only applies to the result(s) stated in the report.

This report superseded any previous report(s) with same work order number.

### EQUIPMENT INFORMATION

Equipment information (Brand name, Model No., Serial No. and Equipment No.) is provided by client.

Equipment Type: Multifunctional Meter

Service Nature: Performance Check

Scope: Conductivity, Dissolved Oxygen, pH Value, Turbidity, Salinity and Temperature

Brand Name/ Model No.: [YSI]/ [ProDSS]

Serial No./ Equipment No.: [24G101660]/ [N/A]

Date of Calibration: 17-October-2025

Ms. Cheng Sin Ying, May  
Senior Chemist - Inorganics

*This report shall not be reproduced except in full without the written approval of the laboratory.*

# REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



**WORK ORDER:** HK2543901  
**SUB-BATCH:** 0  
**DATE OF ISSUE:** 21-Oct-2025  
**CLIENT:** AURECON HONG KONG LIMITED

Equipment Type: Multifunctional Meter  
Brand Name/Model No.: [YSI]/ [ProDSS]  
Serial No.: [24G101660]/ [N/A]  
Equipment No.:  
Date of Calibration: 17-October-2025

Date of Next Calibration: 17-January-2026

## PARAMETERS:

### Conductivity

Method Ref: APHA (23rd edition), 2510B

Expected Reading ( $\mu\text{S}/\text{cm}$ )	Displayed Reading ( $\mu\text{S}/\text{cm}$ )	Tolerance (%)
146.9	147.7	+0.5
6667	6826	+2.4
12890	12959	+0.5
58670	57401	-2.2
	Tolerance Limit (%)	$\pm 10.0$

### Dissolved Oxygen

Method Ref: APHA (23rd edition), 4500O: G

Expected Reading (mg/L)	Displayed Reading (mg/L)	Tolerance (mg/L)
1.58	1.73	+0.15
5.00	4.89	-0.11
6.98	7.07	+0.09
	Tolerance Limit (mg/L)	$\pm 0.20$

### pH Value

Method Ref: APHA (23rd edition), 4500H: B

Expected Reading (pH unit)	Displayed Reading (pH unit)	Tolerance (pH unit)
4.0	3.88	-0.12
7.0	7.09	+0.09
10.0	9.96	-0.04
	Tolerance Limit (pH unit)	$\pm 0.20$

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Cheng Sin Ying, May  
Senior Chemist - Inorganics

# REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



**WORK ORDER:** HK2543901  
**SUB-BATCH:** 0  
**DATE OF ISSUE:** 21-Oct-2025  
**CLIENT:** AURECON HONG KONG LIMITED

Equipment Type: Multifunctional Meter  
Brand Name/Model No.: [YSI]/ [ProDSS]  
Serial No.: [24G101660]/ [N/A]  
Equipment No.:  
Date of Calibration: 17-October-2025

Date of Next Calibration: 17-January-2026

## PARAMETERS:

### Turbidity

Method Ref: APHA (23rd edition), 2130B

Expected Reading (NTU)	Displayed Reading (NTU)	Tolerance (%)
0	0.44	--
4	4.26	+6.5
40	37.62	-6.0
80	76.36	-4.6
400	362.64	-9.3
800	733.83	-8.3
Tolerance Limit (%)		±10.0

### Salinity

Method Ref: APHA (23rd edition), 2520B

Expected Reading (ppt)	Displayed Reading (ppt)	Tolerance (%)
0	0.00	--
10	10.15	+1.5
20	20.56	+2.8
30	29.38	-2.1
Tolerance Limit (%)		±10.0

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Cheng Sin Ying, May  
Senior Chemist - Inorganics

# REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



**WORK ORDER:** HK2543901  
**SUB-BATCH:** 0  
**DATE OF ISSUE:** 21-Oct-2025  
**CLIENT:** AURECON HONG KONG LIMITED

Equipment Type: Multifunctional Meter  
Brand Name/ Model No.: [YSI]/ [ProDSS]  
Serial No./ Equipment No.: [24G101660]/ [N/A]  
Date of Calibration: 17-October-2025

Date of Next Calibration: 17-January-2026

## PARAMETERS:

### Temperature

**Method Ref: Section 6 of International Accreditation New Zealand Technical Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.**

Expected Reading (°C)	Displayed Reading (°C)	Tolerance (°C)
9.0	9.6	+0.6
19.5	18.3	-1.2
39.5	38.3	-1.2
Tolerance Limit (°C)		±2.0

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Cheng Sin Ying, May  
Senior Chemist - Inorganics