



Acumen Laboratory and Testing Limited

Workshop 04, 7/F, The Whitney, No. 183 Wai Yip Street, Kwun Tong, Kowloon

Tel: (852) 2333 6823 Fax: (852) 2333 1316

Page 1 of 1

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
01/11/2025	R252440	0.22	0.10	3.10	3.01	2.95	10	94.1	Pass
03/11/2025	R252460	0.22	0.10	4.29	4.44	-3.44	10	93.1	Pass
05/11/2025	R252481	0.22	0.11	4.79	4.62	3.61	10	93.7	Pass
08/11/2025	R252499	0.22	0.07	4.68	4.85	-3.57	10	92.7	Pass
10/11/2025	R252523	0.22	0.08	4.02	3.91	2.77	10	92.8	Pass
12/11/2025	R252531	0.22	0.09	5.02	5.17	-2.94	10	92.0	Pass
14/11/2025	R252542	0.22	0.10	3.99	3.87	3.05	10	94.3	Pass
17/11/2025	R252556	0.22	0.09	3.01	3.09	-2.62	10	92.5	Pass
19/11/2025	R252582	0.22	0.07	4.14	3.97	4.19	10	92.6	Pass
21/11/2025	R252599	0.22	0.10	4.33	4.42	-2.06	10	93.8	Pass
24/11/2025	R252606	0.22	0.10	4.44	4.30	3.20	10	93.7	Pass
26/11/2025	R252639	0.22	0.07	4.04	4.17	-3.17	10	94.2	Pass
28/11/2025	R252670	0.22	0.09	4.36	4.22	3.26	10	92.0	Pass