



Acumen Laboratory and Testing Limited

Flat/Rm D, 12/F, Ford Glory Plaza, Nos. 37-39 Wing Hong Street, Cheung Sha Wan, Kowloon, Hong Kong
 Tel: (852) 2333 6823 Fax: (852) 2333 1316

Appendix - Quality Control Summary Table

Project Name: Hung Shui Kiu/Ha Tsuen New Development Area Stage 1 Works

		Method Blank Report		Duplicate Report			Sample Spike Report		Pass / Fail
		MDL	Result	Original Result	Duplicate Result	RPD	Spike concentration	Spike Recovery	
Sampling Date	Job No. / Unit	mg/L	mg/L	mg/L	mg/L	%	mg/L	%	/
R231595	03/10/2023	0.22	0.07	4.40	4.52	-2.7	10	92.6	Pass
R231600	05/10/2023	0.22	0.09	6.11	5.83	4.7	10	94.8	Pass
R231601	10/10/2023	0.22	0.08	5.19	4.91	5.5	10	93.7	Pass
R231625	12/10/2023	0.22	0.07	5.65	5.89	-4.2	10	91.1	Pass
R231626	14/10/2023	0.22	0.07	4.87	5.08	-4.2	10	94.6	Pass
R231627	17/10/2023	0.22	0.08	4.99	4.77	4.5	10	92.8	Pass
R231675	19/10/2023	0.22	0.06	5.42	5.56	-2.6	10	94.4	Pass
R231697	21/10/2023	0.22	0.07	4.05	4.26	-5.0	10	91.6	Pass
R231698	24/10/2023	0.22	0.06	4.30	4.03	6.5	10	91.0	Pass
R231760	26/10/2023	0.22	0.09	5.04	4.81	4.7	10	92.7	Pass
R231761	28/10/2023	0.22	0.07	4.25	4.54	-6.6	10	92.9	Pass
R231762	31/10/2023	0.22	0.10	4.92	5.19	-5.3	10	93.0	Pass